

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.1	0.09916	0.840	± 1.2	±	3.2E-01 mm						
0 a 0.5	0.49917	0.830	± 1.6	±	3.2E-01 mm						
0 a 1	0.99815	1.850	± 2	±	3.2E-01 mm						
0 a 2	1.99781	2.190	± 2.8	±	3.9E-01 mm						
0 a 3	2.99788	2.120	± 3.6	±	3.9E-01 mm						
0 a 4	3.99767	2.330	± 4.4	±	3.9E-01 mm						
0 a 5	4.99724	2.760	± 5.2	±	3.9E-01 mm						
0 a 6	5.99685	3.150	± 6	±	2.1E-01 mm						
0 a 7	6.99602	3.980	± 6.8	±	2.1E-01 mm						
0 a 8	7.99574	4.260	± 7.6	±	2.1E-01 mm						