

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.131 °C	0.0 °C	0.131 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.085 °C	20.2 °C	0.115 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.585 °C	34.9 °C	0.315 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.159 %HR	31.7 %HR	2.541 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	54.343 %HR	54.4 %HR	0.057 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.427 %HR	80.0 %HR	-0.427 %HR	± 3.0 %HR	± 1.9E+00 %HR