

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.131 °C	0.1 °C	0.231 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.085 °C	20.3 °C	0.215 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.585 °C	34.5 °C	-0.085 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.159 %HR	31.8 %HR	2.641 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	54.343 %HR	54.2 %HR	-0.143 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.427 %HR	81.3 %HR	0.873 %HR	± 3.0 %HR	± 1.9E+00 %HR