

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 0 °C              | -0.131 °C         | 0.0 °C    | 0.131 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 20 °C             | 20.085 °C         | 20.2 °C   | 0.115 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 35 °C             | 34.585 °C         | 34.9 °C   | 0.315 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 29.159 %HR        | 30.6 %HR  | 1.441 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 55 %HR            | 54.343 %HR        | 55.9 %HR  | 1.557 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 80 %HR            | 80.427 %HR        | 83.1 %HR  | 2.673 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |