

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
15 °C	15.181 °C	15.3 °C	0.119 °C	±	1 °C	± 4.0E-01 °C
30 °C	29.541 °C	29.7 °C	0.159 °C	±	1 °C	± 4.0E-01 °C
35 °C	35.156 °C	35.3 °C	0.144 °C	±	1 °C	± 4.0E-01 °C
40 %HR	38.903 %HR	41.4 %HR	2.497 %HR	±	5 %HR	± 1.9E+00 %HR
50 %HR	50.294 %HR	52.2 %HR	1.906 %HR	±	5 %HR	± 1.9E+00 %HR
90 %HR	89.691 %HR	88.1 %HR	-1.591 %HR	±	5 %HR	± 1.9E+00 %HR