

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.5 mm	0.50 mm	0 µm	± 10 µm	±	5.8E+00 µm
1 mm	1.01 mm	10 µm	± 10 µm	±	5.8E+00 µm
2 mm	2.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
3 mm	3.01 mm	10 µm	± 10 µm	±	5.8E+00 µm
35 mm	35.01 mm	10 µm	± 10 µm	±	5.8E+00 µm
40 mm	40.01 mm	10 µm	± 10 µm	±	5.8E+00 µm
45 mm	45.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
50 mm	50.01 mm	10 µm	± 10 µm	±	5.8E+00 µm