

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00 mm	1.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
1.25 mm	1.25 mm	0 µm	± 10 µm	±	5.8E+00 µm
1.50 mm	1.50 mm	0 µm	± 10 µm	±	5.8E+00 µm
2.00 mm	2.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
3.00 mm	3.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
4.00 mm	4.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
8.00 mm	8.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
10.00 mm	10.00 mm	0 µm	± 10 µm	±	5.8E+00 µm