

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
23 °C	23.043 °C	23.2 °C	0.157 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.163 °C	30 °C	-0.063 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.775 %HR	32 %HR	1.025 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.222 %HR	52 %HR	1.478 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.867 %HR	76.7 %HR	0.833 %HR	± 3.0 %HR	± 1.9E+00 %HR