

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
23.3 μm	24 μm	0.2 μm	± 0.73 μm	±	2.9E-01 μm
51.6 μm	52 μm	-0.1 μm	± 1.0 μm	±	2.9E-01 μm
124.6 μm	124 μm	-0.6 μm	± 5.7 μm	±	2.9E-01 μm
250.3 μm	250 μm	-0.3 μm	± 9.5 μm	±	2.9E-01 μm
486.0 μm	486 μm	0.0 μm	± 17 μm	±	3.0E-01 μm