

Patrón (Standard)		UBP (UUT)		Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.00	mm	1.000	mm	0 µm	± 6 µm	± 5.9E-01 µm
1.25	mm	1.250	mm	0 µm	± 4 µm	± 5.8E-01 µm
1.50	mm	1.500	mm	0 µm	± 4 µm	± 5.8E-01 µm
2.00	mm	2.000	mm	0 µm	± 4 µm	± 5.8E-01 µm
3.00	mm	3.000	mm	0 µm	± 4 µm	± 5.8E-01 µm
5.00	mm	5.000	mm	0 µm	± 4 µm	± 5.8E-01 µm
8.00	mm	8.000	mm	0 µm	± 4 µm	± 5.9E-01 µm
10.00	mm	10.000	mm	0 µm	± 4 µm	± 5.9E-01 µm
15.00	mm	15.000	mm	0 µm	± 4 µm	± 5.8E-01 µm
20.00	mm	20.001	mm	1 µm	± 4 µm	± 5.9E-01 µm
25.00	mm	25.002	mm	2 µm	± 4 µm	± 5.9E-01 µm