

Patrón (Standard)		UBP (UUT)		Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.00	mm	1.000	mm	0 μm	± 6 μm	±	5.9E-01 μm
1.25	mm	1.250	mm	0 μm	± 4 μm	±	5.8E-01 μm
1.50	mm	1.500	mm	0 μm	± 4 μm	±	5.8E-01 μm
2.00	mm	2.000	mm	0 μm	± 4 μm	±	5.8E-01 μm
3.00	mm	3.000	mm	0 μm	± 4 μm	±	5.8E-01 μm
5.00	mm	5.000	mm	0 μm	± 4 μm	±	5.8E-01 μm
8.00	mm	8.000	mm	0 μm	± 4 μm	±	5.9E-01 μm
10.00	mm	10.001	mm	1 μm	± 4 μm	±	5.9E-01 μm
15.00	mm	15.001	mm	1 μm	± 4 μm	±	5.8E-01 μm
20.00	mm	20.001	mm	1 μm	± 4 μm	±	5.9E-01 μm
25.00	mm	25.001	mm	1 μm	± 4 μm	±	5.9E-01 μm