

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	0.99953	0.470	± 1	±	4.3E-01 mm						
0 a 2	1.99958	0.420	± 1.4	±	4.8E-01 mm						
0 a 3	2.99918	0.820	± 1.8	±	4.8E-01 mm						
0 a 4	3.99918	0.820	± 2.2	±	4.8E-01 mm						
0 a 5	4.99922	0.780	± 2.6	±	4.8E-01 mm						
0 a 6	5.99954	0.460	± 3	±	3.6E-01 mm						
0 a 7	6.99937	0.630	± 3.4	±	3.6E-01 mm						
0 a 8	7.99962	0.380	± 3.8	±	3.6E-01 mm						
0 a 9	8.99913	0.870	± 4.2	±	3.6E-01 mm						
0 a 10	9.99915	0.850	± 4.6	±	3.6E-01 mm						