

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
2 μm	2.02 μm	0.02 μm	± 2.5 μm	±	1.1E-01 μm
5 μm	4.97 μm	-0.03 μm	± 2.5 μm	±	9.6E-02 μm
15 μm	14.87 μm	-0.13 μm	± 2.5 μm	±	9.0E-02 μm
20 μm	19.76 μm	-0.24 μm	± 2.5 μm	±	1.2E-01 μm
25 μm	24.68 μm	-0.32 μm	± 2.5 μm	±	1.4E-01 μm