

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49890	1.100	± 1.6	± 4.3E-01 mm					
0 a 1	0.99889	1.110	± 2	± 4.3E-01 mm					
0 a 1.5	1.49918	0.820	± 2.4	± 4.8E-01 mm					
0 a 2	1.99919	0.810	± 2.8	± 4.8E-01 mm					
0 a 2.5	2.49916	0.840	± 3.2	± 4.8E-01 mm					
0 a 3	2.99917	0.830	± 3.6	± 4.8E-01 mm					
0 a 3.5	3.49929	0.710	± 4	± 4.8E-01 mm					
0 a 4	3.99925	0.750	± 4.4	± 4.8E-01 mm					
0 a 4.5	4.49916	0.840	± 4.8	± 4.8E-01 mm					
0 a 5	0.89000	0.890	± 5.2	± 4.8E-01 mm					