

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.5	0.49941	0.59	± 0.8	±	4.3E-01 mm						
0 a 1	0.99936	0.64	± 1	±	4.3E-01 mm						
0 a 1.5	1.49960	0.40	± 1.2	±	4.8E-01 mm						
0 a 2	1.99956	0.44	± 1.4	±	4.8E-01 mm						
0 a 2.5	2.49948	0.52	± 1.6	±	4.8E-01 mm						
0 a 3	2.99943	0.57	± 1.8	±	4.8E-01 mm						
0 a 3.5	3.49947	0.53	± 2	±	4.8E-01 mm						
0 a 4	3.99936	0.64	± 2.2	±	4.8E-01 mm						
0 a 4.5	4.49919	0.81	± 2.4	±	4.8E-01 mm						
0 a 5	4.99906	0.94	± 2.6	±	4.8E-01 mm						