

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.1	0.09918	0.820	± 1.2	±	4.3E-01 mm						
0 a 0.3	0.29915	0.850	± 1.4	±	4.3E-01 mm						
0 a 0.4	0.39912	0.880	± 1.5	±	4.3E-01 mm						
0 a 0.5	0.49909	0.910	± 1.6	±	4.3E-01 mm						
0 a 0.7	0.69908	0.920	± 1.7	±	4.3E-01 mm						
0 a 0.8	0.79907	0.930	± 1.8	±	4.3E-01 mm						
0 a 0.9	0.89907	0.930	± 1.9	±	4.3E-01 mm						
0 a 1	0.99902	0.980	± 2	±	4.3E-01 mm						
0 a 1.5	1.49889	1.110	± 2.4	±	4.8E-01 mm						
0 a 2	1.09000	1.090	± 2.8	±	4.8E-01 mm						