

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09918	0.820	± 1.2	± 4.3E-01 mm					
0 a 0.3	0.29915	0.850	± 1.4	± 4.3E-01 mm					
0 a 0.4	0.39912	0.880	± 1.5	± 4.3E-01 mm					
0 a 0.5	0.49909	0.910	± 1.6	± 4.3E-01 mm					
0 a 0.7	0.69908	0.920	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.79907	0.930	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.89907	0.930	± 1.9	± 4.3E-01 mm					
0 a 1	0.99902	0.980	± 2	± 4.3E-01 mm					
0 a 1.5	1.49889	1.110	± 2.4	± 4.8E-01 mm					
0 a 2	1.99891	1.090	± 2.8	± 4.8E-01 mm					