

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00 mm	1.000 mm	0 µm	± 30 µm	±	5.8E+01 µm
1.50 mm	1.500 mm	0 µm	± 30 µm	±	5.8E+01 µm
2.00 mm	1.980 mm	-20 µm	± 30 µm	±	5.8E+01 µm
3.00 mm	2.980 mm	-20 µm	± 30 µm	±	5.8E+01 µm
5.00 mm	4.980 mm	-20 µm	± 30 µm	±	5.8E+01 µm
10.00 mm	9.970 mm	-30 µm	± 30 µm	±	5.8E+01 µm
15.00 mm	14.970 mm	-30 µm	± 30 µm	±	5.8E+01 µm
20.00 mm	19.970 mm	-30 µm	± 30 µm	±	5.8E+01 µm
25.00 mm	24.970 mm	-30 µm	± 30 µm	±	5.8E+01 µm