

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm	± 30 µm	±	5.8E+01 µm
1.50 mm	1.50 mm	0 µm	± 30 µm	±	5.8E+01 µm
2.00 mm	1.98 mm	-20 µm	± 30 µm	±	5.8E+01 µm
3.00 mm	2.98 mm	-20 µm	± 30 µm	±	5.8E+01 µm
5.00 mm	4.98 mm	-20 µm	± 30 µm	±	5.8E+01 µm
10.00 mm	9.97 mm	-30 µm	± 30 µm	±	5.8E+01 µm
15.00 mm	14.97 mm	-30 µm	± 30 µm	±	5.8E+01 µm
20.00 mm	19.97 mm	-30 µm	± 30 µm	±	5.8E+01 µm
25.00 mm	24.97 mm	-30 µm	± 30 µm	±	5.8E+01 µm