

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1 mm	0.999 mm	-1 µm	± 4 µm	±	5.8E-01 µm
1.25 mm	1.252 mm	2 µm	± 4 µm	±	5.8E-01 µm
1.5 mm	1.500 mm	-1 µm	± 4 µm	±	5.8E-01 µm
2 mm	1.999 mm	-1 µm	± 4 µm	±	5.8E-01 µm
3 mm	3.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
4 mm	4.000 mm	0 µm	± 4 µm	±	5.9E-01 µm
5 mm	4.999 mm	-1 µm	± 4 µm	±	5.8E-01 µm
10 mm	9.999 mm	-1 µm	± 4 µm	±	5.8E-01 µm
15 mm	15.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
20 mm	20.002 mm	2 µm	± 4 µm	±	5.9E-01 µm
25 mm	25.001 mm	1 µm	± 4 µm	±	5.9E-01 µm