

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	19.965 °C	19.8 °C	-0.165 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.9 °C	0.350 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.0 °C	-0.363 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	43.6 %HR	13.337 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.652 %HR	62.4 %HR	11.748 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	76.1 %HR	5.959 %HR	± 3.0 %HR	± 1.9E+00 %HR