

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	19.965 °C	19.9 °C	-0.065 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.9 °C	0.350 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	29.9 °C	-0.463 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	43.1 %HR	12.837 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.652 %HR	62.2 %HR	11.548 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	76.9 %HR	6.759 %HR	± 3.0 %HR	± 1.9E+00 %HR