

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.06 in	-0.003 in	± 0.03 in	±	5.8E-03 in
0.100 in	0.10 in	0.000 in	± 0.03 in	±	5.8E-03 in
0.125 in	0.12 in	-0.005 in	± 0.03 in	±	5.8E-03 in
0.200 in	0.20 in	0.000 in	± 0.03 in	±	5.8E-03 in
0.250 in	0.25 in	0.000 in	± 0.03 in	±	5.8E-03 in
0.300 in	0.30 in	0.000 in	± 0.03 in	±	5.8E-03 in
0.400 in	0.40 in	0.000 in	± 0.03 in	±	5.8E-03 in
0.450 in	0.45 in	0.000 in	± 0.03 in	±	5.8E-03 in
0.500 in	0.50 in	0.000 in	± 0.03 in	±	5.8E-03 in