

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C             | 9.957 °C          | 8.5 °C    | -1.457 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 15 °C             | 14.921 °C         | 13 °C     | -1.921 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.049 °C         | 23 °C     | -1.749 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 32 °C             | 32.369 °C         | 32 °C     | -0.669 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.570 %HR        | 33.7 %HR  | 3.130 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 70.7 %HR  | 0.252 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |