

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
7 °C	7.199 °C	7.6 °C	0.401 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.025 °C	20 °C	0.275 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	32 °C	-0.368 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33 %HR	2.030 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.263 %HR	54.9 %HR	-0.363 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	79.3 %HR	-1.373 %HR	± 3.0 %HR	± 1.9E+00 %HR