

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
7 °C	7.199 °C	7.2 °C	0.001 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.025 °C	20 °C	-0.125 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	32 °C	0.032 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33 %HR	2.430 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.263 %HR	58.0 %HR	2.737 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	78.7 %HR	-1.973 %HR	± 3.0 %HR	± 1.9E+00 %HR