

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 7 °C              | 7.199 °C          | 7.3 °C    | 0.101 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 20 °C             | 20.025 °C         | 20 °C     | -0.125 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 32 °C             | 32.068 °C         | 32 °C     | 0.232 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.570 %HR        | 31 %HR    | 0.330 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 55 %HR            | 55.263 %HR        | 55.6 %HR  | 0.337 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 80 %HR            | 80.673 %HR        | 80.4 %HR  | -0.273 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |