

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.363 °C	31.5 °C	1.137 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	- %HR	- %HR	± 5.0 %HR	± - %HR
50 %HR	50.376 %HR	36.8 %HR	-13.576 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.754 %HR	63.5 %HR	-7.254 %HR	± 5.0 %HR	± 1.9E+00 %HR