

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.862 °C	29.6 °C	-0.262 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	30.0 %HR	-0.059 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.805 %HR	49.0 %HR	-1.805 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	73.0 %HR	2.859 %HR	± 5.0 %HR	± 1.9E+00 %HR