

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
cm	cm	in	in	(Uncertainty)	cm	cm	in	in	(Uncertainty)
0 a 10	9.99817	0.022	± 0.18	± 3.6E-04 in					
0 a 20	19.99455	0.065	± 4.2	± 3.6E-04 in					
0 a 30	29.98853	0.138	± 6.2	± 5.5E-04 in					
0 a 40	39.98023	0.237	± 8.2	± 5.5E-04 in					
0 a 50	49.96972	0.363	± 10.2	± 6.3E-04 in					
0 a 60	59.95676	0.519	± 12.2	± 6.3E-04 in					
0 a 70	69.94150	0.702	± 14.2	± 7.1E-04 in					
0 a 80	79.92394	0.913	± 16.2	± 7.1E-04 in					
0 a 90	89.90402	1.152	± 18.2	± 5.7E-04 in					
0 a 100	99.88180	1.418	± 20.2	± 5.7E-04 in					
0 a 110	109.99981	1.712	± 22.2	± 5.7E-04 in					
0 a 120	119.99969	2.036	± 24.2	± 5.7E-04 in					