

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.1	0.09949	0.51	± 0.6	±	4.3E-01 mm						
0 a 0.2	0.19953	0.47	± 0.6	±	4.3E-01 mm						
0 a 0.3	0.29947	0.53	± 0.7	±	4.3E-01 mm						
0 a 0.4	0.39946	0.54	± 0.7	±	4.3E-01 mm						
0 a 0.5	0.49955	0.45	± 0.8	±	4.3E-01 mm						
0 a 0.6	0.59952	0.48	± 0.8	±	4.3E-01 mm						
0 a 0.7	0.69952	0.48	± 0.8	±	4.3E-01 mm						
0 a 0.8	0.79952	0.48	± 0.9	±	4.3E-01 mm						
0 a 0.9	0.89949	0.51	± 0.9	±	4.3E-01 mm						
0 a 1	0.99947	0.53	± 1	±	4.3E-01 mm						
0 a 1.1	1.09952	0.48	± 1	±	4.8E-01 mm						
0 a 1.2	1.19940	0.60	± 1	±	4.8E-01 mm						