

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.322 °C	14.6 °C	-0.722 °C	± 0.3 °C	± 4.0E-01 °C
20 °C	20.035 °C	19.3 °C	-0.735 °C	± 0.3 °C	± 4.0E-01 °C
25 °C	25.650 °C	24.8 °C	-0.850 °C	± 0.3 °C	± 4.0E-01 °C
30 °C	30.263 °C	29.8 °C	-0.463 °C	± 0.3 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	27.4 %HR	-2.761 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.294 %HR	47.3 %HR	-2.994 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	77.9 %HR	-2.773 %HR	± 3.0 %HR	± 1.9E+00 %HR