

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.286 °C	3.0 °C	0.714 °C	± 1.0 °C	± 4.0E-01 °C
5 °C	4.993 °C	6.0 °C	1.007 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.002 °C	9.0 °C	0.998 °C	± 1.0 °C	± 4.0E-01 °C