

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.322 °C	16.0 °C	0.678 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.035 °C	21.0 °C	0.965 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.650 °C	27.0 °C	1.350 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.263 °C	31.0 °C	0.737 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	39.889 °C	41.0 °C	1.111 °C	± 1.0 °C	± 4.0E-01 °C