

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.308 °C	11.0 °C	0.692 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.035 °C	21.0 °C	0.965 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	40.090 °C	41.0 °C	0.910 °C	± 1.0 °C	± 4.0E-01 °C