

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.3	0.29993	0.070	± 0.2	±	4.3E-01 mm						
0 a 0.6	0.59991	0.090	± 0.3	±	4.3E-01 mm						
0 a 0.9	0.89989	0.110	± 0.3	±	4.3E-01 mm						
0 a 1.2	1.19989	0.110	± 0.4	±	4.8E-01 mm						
0 a 1.5	1.50005	-0.050	± 0.5	±	4.8E-01 mm						
0 a 1.8	1.79985	0.150	± 0.5	±	4.8E-01 mm						
0 a 2.1	2.09986	0.140	± 0.6	±	4.8E-01 mm						
0 a 2.4	2.39957	0.430	± 0.6	±	4.8E-01 mm						
0 a 2.7	2.69946	0.540	± 0.7	±	4.8E-01 mm						
0 a 3	2.99937	0.630	± 0.8	±	4.8E-01 mm						