

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.3	0.29993	0.070	± 0.2	± 4.3E-01 mm					
0 a 0.6	0.59995	0.050	± 0.3	± 4.3E-01 mm					
0 a 0.9	0.89989	0.110	± 0.3	± 4.3E-01 mm					
0 a 1.2	1.19986	0.140	± 0.4	± 4.8E-01 mm					
0 a 1.5	1.50003	-0.030	± 0.5	± 4.8E-01 mm					
0 a 1.8	1.79980	0.200	± 0.5	± 4.8E-01 mm					
0 a 2.1	2.10001	-0.010	± 0.6	± 4.8E-01 mm					
0 a 2.4	2.39956	0.440	± 0.6	± 4.8E-01 mm					
0 a 2.7	2.69944	0.560	± 0.7	± 4.8E-01 mm					
0 a 3	2.99928	0.720	± 0.8	± 4.8E-01 mm					