

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.021 °C	15.3 °C	0.279 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.3 °C	-0.163 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	23.0 %HR	2.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	76.0 %HR	-4.673 %HR	± 5.0 %HR	± 1.9E+00 %HR