

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.021 °C	14.6 °C	-0.421 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.1 °C	-0.363 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	24.0 %HR	3.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	61.0 %HR	-4.335 %HR	± 5.0 %HR	± 1.9E+00 %HR