

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
15 °C	15.021 °C	14.4 °C	-0.621 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.463 °C	29.9 °C	-0.563 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	25.0 %HR	4.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	78.0 %HR	-2.673 %HR	± 5.0 %HR	± 1.9E+00 %HR