

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.021 °C	13.7 °C	-1.321 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.7 °C	0.237 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	27.0 %HR	6.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	74.0 %HR	8.665 %HR	± 5.0 %HR	± 1.9E+00 %HR