

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10 RPM	10 RPM	0 RPM	± 0.11 RPM	±	5.8E-02 RPM
60 RPM	60 RPM	0 RPM	± 0.16 RPM	±	5.8E-02 RPM
400 RPM	400 RPM	0 RPM	± 0.50 RPM	±	5.8E-02 RPM
1000 RPM	1000 RPM	0 RPM	± 2.0 RPM	±	5.8E-01 RPM
10000 RPM	10000 RPM	0 RPM	± 6.0 RPM	±	8.0E-01 RPM