

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.263 °C	30.1 °C	-0.163 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	32.4 %HR	2.341 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.294 %HR	50.8 %HR	0.506 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.0 %HR	0.552 %HR	± 3.0 %HR	± 1.9E+00 %HR