

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.263 °C	30.3 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	32.0 %HR	1.941 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.294 %HR	49.6 %HR	-0.694 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	72.1 %HR	1.652 %HR	± 3.0 %HR	± 1.9E+00 %HR