

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 1 mm                     | 1.00 mm          | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |
| 3 mm                     | 3.00 mm          | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |
| 50 mm                    | 50.00 mm         | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |
| 60 mm                    | 60.00 mm         | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |
| 70 mm                    | 70.00 mm         | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |
| 80 mm                    | 80.00 mm         | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |
| 90 mm                    | 90.00 mm         | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |
| 100 mm                   | 100.00 mm        | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |
| 105 mm                   | 105.00 mm        | 0 μm                 | ± 30 μm                     | ±                                  | 5.8E+00 μm |