

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.05026	-0.260	± 0.016	± 4.3E-01 mm					
0 a 0.1	0.10025	-0.250	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.15023	-0.230	± 0.6	± 4.3E-01 mm					
0 a 1	1.00016	-0.160	± 0.6	± 4.3E-01 mm					
0 a 2	2.00008	-0.080	± 1	± 4.8E-01 mm					
0 a 3	2.99923	0.770	± 1.4	± 4.8E-01 mm					
0 a 4	3.99928	0.720	± 1.8	± 4.8E-01 mm					
0 a 5	5.00009	-0.090	± 2.2	± 4.8E-01 mm					
0 a 6	6.00009	-0.090	± 2.6	± 3.6E-01 mm					
0 a 7	6.99936	0.640	± 3	± 3.6E-01 mm					
0 a 8	7.99929	0.710	± 3.4	± 3.6E-01 mm					
0 a 9	8.99909	0.910	± 3.8	± 3.6E-01 mm					
0 a 10	9.99826	1.740	± 4.2	± 3.6E-01 mm					
0 a 11	10.99911	0.890	± 4.6	± 3.6E-01 mm					
0 a 12	12.00016	-0.160	± 5	± 3.6E-01 mm					
0 a 13	12.99966	0.340	± 5.4	± 3.6E-01 mm					
0 a 14	14.00023	-0.230	± 5.8	± 3.6E-01 mm					
0 a 15	14.99972	0.280	± 6.2	± 3.6E-01 mm					
0 a 16	15.99974	0.260	± 6.6	± 3.6E-01 mm					
0 a 17	16.99841	1.590	± 7	± 3.6E-01 mm					
0 a 18	17.99940	0.600	± 7.4	± 3.6E-01 mm					
0 a 19	18.99973	0.270	± 7.8	± 3.6E-01 mm					
0 a 20	20.00053	-0.530	± 8.2	± 3.6E-01 mm					