

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.308 °C	10.3 °C	-0.008 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.0 °C	-0.021 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.1 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	31.9 °C	-0.168 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	31.0 %HR	0.839 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.754 %HR	69.7 %HR	-1.054 %HR	± 3.0 %HR	± 1.9E+00 %HR