

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.293 °C	30.5 °C	0.207 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.486 %HR	32.3 %HR	2.814 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.621 %HR	52.9 %HR	2.279 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.499 %HR	73.7 %HR	-1.799 %HR	± 3.0 %HR	± 1.9E+00 %HR