

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.293 °C	30.2 °C	-0.093 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.486 %HR	32.4 %HR	2.914 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.621 %HR	51.2 %HR	0.579 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.499 %HR	72.7 %HR	-2.799 %HR	± 3.0 %HR	± 1.9E+00 %HR