

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	30.2 °C	-0.093 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.486 %HR	32.4 %HR	2.914 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.621 %HR	51.1 %HR	0.479 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.499 %HR	73.4 %HR	-2.099 %HR	± 3.0 %HR	± 1.9E+00 %HR